



Policy

Saskatoon & Area - Occupational Health & Safety

Number: 55-005

Title: Other Hazards: Ethylene Oxide Sterilizers

Saskatchewan Employment Act:

OHS Regulation: 477

Date: January 1, 2017

Date Revised/Reaffirmed: June 4, 2018

Any PRINTED version of this document is only accurate up to the date of printing. Saskatchewan Health Authority (SHA), Occupational Health & Safety Services (OHS) cannot guarantee the currency or accuracy of any printed policy. Always refer to the Occupational Health & Safety internal website for the most current versions of documents in effect. This policy complies with legislation (minimum standards) at the time of this writing and/or revision. SHA OHS accepts no responsibility for use of this material by any person or organization not associated with SHA. No part of this document may be reproduced in any form for publication without permission of SHA OHS Services.

Purpose

To ensure that ethylene oxide sterilizers are operated and maintained in accordance with the CSA installation standard

Definition

- **“CSA installation standard”** means the Canadian Standards Association standard CAN/CSA-Z314.9-M89 Installation and Ventilation of Ethylene Oxide Sterilizers in Health Care Facilities

Roles and Responsibilities

Employer must ensure:

- That to the extent that is practicable that all ethylene oxide sterilizers are operated and maintained according with the CSA standard
- That they work with the OHC committee to develop safe work procedures and policies
- That the work with OHC to prepare an emergency response program
- That any ethylene oxide sterilizers purchased after these regulations came into force:
 - Is constructed in accordance with the CSA standard CAN/CSA-Z314.1-M91
 - Is installed and meets the ventilation requirements of the CSA standard
 - Where reasonably practicable, is a sterilizer with in-chamber aeration that allows sterilization and aeration to take place without manually transferring the items that are being sterilized and aerated from one piece of equipment to another.
- An employer shall ensure that portable ethylene oxide sterilizers are operated in a fume cabinet or placed in a self-contained room that is unoccupied during the sterilization process and is ventilated clear of the place of employment at a minimum rate of 10 air changes per hour to prevent the accumulation of the gas in the room.

Manager/Supervisor must ensure:

- That roles and responsibilities of all workers are being met, as well as applicable legislation and/or standards
- Implement the safe work practices and policies
- All workers who operate ethylene oxide sterilizers or may come in contact with ethylene oxide are trained in accordance with CSA standards
- That where ethylene oxide is used or stored is clearly marked with legible signs that state “ETHYLENE OXIDE AREA, POTENTIAL CANCER AND REPRODUCTIVE HAZARD, AUTHORIZED PERSONNEL ONLY”

- That all records pertaining to equipment maintenance and accidental leaks of ethylene oxide are kept in a log book for 5 years

Worker must ensure:

- They are trained in and follow the safe work procedures
- They are trained in and use any PPE required (selection, use, care and handling)
- They report to their supervisor and the Safety Alert System/Incident Report Line any incidents or near misses

Orientation & Training

- In all safe work procedures and policies regarding ethylene oxide
- In the handling, storage, use and emergency response regarding ethylene oxide or sterilization machines

Materials

- Appropriate PPE if required
- Log books to record accidental spills and maintenance records

Non-Compliance/Breach:

Non-compliance with this policy will result in a review of the incident. A review for non-compliance may result in disciplinary action, up to and including termination of employment or privileges; fines and /or prosecution of individuals under the Saskatchewan Employment Act and OHS Regulations.

References

- CSA Standards for Ethylene Oxide Sterilizers

Review Dates:

January 1, 2017

June 4, 2018